



# **BACHARACH'S NEW STANDARD IN COMBUSTION & EMISSIONS ANALYSIS**



## **Graphic Display**

- 160 x 160 graphic liquid crystal display
- Easy to read
- Backlit
- Zoom

## PCA<sup>®</sup> 2

The PCA® 2 is a commercial grade, handheld combustion and emissions analyzer for on demand or semi-continuous sampling of light industrial, institutional, commercial and residential furnaces, boilers and appliances. The PCA® 2 is the perfect tool for service technicians and boiler contractors who need to ensure safe operating conditions, determine combustion efficiency or perform emissions testing in combustion applications.

The PCA<sup>®</sup> 2 directly measures and displays flue gas Oxygen (O<sub>2</sub>), Carbon Monoxide (CO), Stack Temperature, Draft, Differential Pressure, Combustion Air Temperature and optionally measures and displays Nitric Oxide (NO), Nitrogen Dioxide (NO<sub>2</sub>) and Sulfur Dioxide (SO<sub>2</sub>). Simultaneously, the PCA<sup>®</sup>2 calculates and displays Combustion Efficiency (EFF), Excess Air (EA), Carbon Dioxide (CO<sub>2</sub>), NOx and Oxygen reference values. The PCA<sup>®</sup> 2 performs combustion calculations for ten fuels including, Natural Gas, Oil #2, Oil #4, Oil #6, Propane, Coal, Wood, Kerosene, Bagasse and Digester Gas. The large backlit graphic display shows eight different measurements and calculated values simultaneously, and also has zoom capabilities.



SPECIFICATION	1S	PCA <sup>®</sup> 2	PCA <sup>®</sup> 2 REPLACEMENT PARTS			
Measurement Ranges	Primary/Ambient Air Temperature	-4° to 999° F	24-0788	Replacement O <sub>2</sub> sensor		
	Stack Temperature Oxygen	-4° to 2192° F 0 to 20.9%	24-0789	Replacement CO sensor		
	Carbon Monoxide (H <sub>2</sub> comp)	0 to 4,000 ppm	24-0789	Replacement NO sensor		
	Carbon Monoxide, high range	4,001 to 20,000 ppm	24-0001	Replacement CO (high range)		
	Nitric Oxide Nitrogen Dioxide	0 to 3,000 ppm 0 to 500 ppm				
	Sulfur Dioxide	0 to 5,000 ppm	24-0998	Replacement SO <sub>2</sub> sensor		
	Pressure/Draft	-72 to +72 inwc	24-1027	Replacement NO <sub>2</sub> sensor		
Calculated Ranges	Combustion Efficiency Excess Air	0.1 to 100% 1.0 to 250%	24-1395	Smart sensor, CO		
	Carbon Dioxide (dry basis)	0 to fuel dependent maximum	24-1397	Smart sensor, CO (high range)		
	$NOx (NOx = NO + NO_2)$	0 to 3,500 ppm	24-1398	Smart sensor, SO <sub>2</sub>		
	NOx referenced to % O <sub>2</sub> CO referenced to % O <sub>2</sub>	0 to 9,999 ppm 0 to 9,999 ppm	24-1399	Smart sensor, NO <sub>2</sub>		
	NO referenced to $\% O_2$	0 to 9,999 ppm	24-1401	Smart sensor, NO		
	$NO_2$ referenced to % $\tilde{O}_2$	0 to 9,999 ppm	204-0020	Battery, NO bias battery		
	SO <sub>2</sub> referenced to % O <sub>2</sub>	0 to 9,999 ppm	24-7059	Calibration kit (less calibration gas)		
Accuracy	Oxygen	$\pm$ 0.3% O <sub>2 (on flue gas)</sub>	24-1400	IrDA Printer with disposable batteries		
	Stack or Flue Gas Temperature	$\pm$ 4°F between 32 and 255°F	24-1310	Printer paper, 5 rolls		
		( $\pm$ 2°C between 0 and 124°C)	06-8733	Printer paper, 1 roll		
		$\pm$ 6°F between 257 and 480°F	24-3004	Probe assembly		
		$(\pm 3^{\circ}\text{C} \text{ between 125 and 249}^{\circ}\text{C})$	19-3265	Replacement water trap assembly		
			07-1644	Replacement filter element (pkg of 3)		
		$\pm$ 8°F between 482 and 752°F ( $\pm$ 4°C between 250 and 400°C)	24-8414	Replacement thermocouple, 12 inch		
			19-3037	Probe stop		
	Primary-air/Ambient Temperature Pressure/Draft	$\pm$ 2°F between 32 and 212°F ( $\pm$ 1°C between 0 and 100°C)	24-1124	Extended hose assembly, 20 ft.		
			104-1797	Thermocouple, 10 ft. (combustion air temp.)		
		$\pm$ 2% of reading or $\pm$ .02 inwc whichever is greater between 0 and $\pm$ 10 inwc	104-1737	Thermocouple, 1 in. (ambient air temp.)		
		$\pm 3\%$ of reading between $\pm 10$ and $\pm 72$ inwc	24-1409	Protective rubber boot		
	CO	$\pm$ 5% of reading or $\pm$ 10 ppm whichever is greater between 0-2000 ppm CO; $\pm$ 10% of reading between 2001 to 20,000 ppm CO	24-1409			
				AC power adapter		
			104-4032	USB cable		
		$\pm$ 5% of reading or $\pm$ 5 ppm whichever is greater between 0-2000 ppm NO	24-1425	PCA2 Data Recovery Software		
	NO <sub>2</sub> *	$\pm$ 5% of reading or $\pm$ 5 ppm whichever is greater between 0-500 ppm NO $_2$	21-7006	Tru Spot Smoke Tester		
	NO <sub>2</sub>		* 24-7224	Compact Sample Conditioner		
	SO <sub>2</sub> ·	$\pm$ 5% of reading or $\pm$ 10 ppm whichever is greater between 0-2000 ppm $\mathrm{SO}_2$	Instrument (	Comes Complete With:		
Memory	500 complete combustion test records			Hard Carry Case Hard Carry Case Probe and Hose Assembly Factory Calibrated and Installed Sensors Instuction Manual		
Logged Memory	500 complete logged combustion test records					
Selectable Fuels	Natural Gas, Oil #2, Oil #4, Oil #6, Propane, Coal, Wood, Kerosene, Bagasse and Digester Gas					
Size	9"H x 3"W x 2.5"D (22.9 cm x 7.6 cm x 6.3 cm)					
Weight	1.4 lbs (0.6 kg)			Batteries		
Power Source	Four disposable AA alkaline batteri Optional AC Power Adapter	es or NiMH rechargeable batteries providing 10 hours of continuous operation.	Product Bo			
Display	160 x 160 (2.5 in x 2.5 in.) Graphic	USB Cable				
Warm Up Time	60 second total warm up time (Ser	PC Softwa				
			2 Year Wa	irranty		

PCA <sup>®</sup> 2 ORDERING INFORMATION										
MODEL NUMBER	PART NUMBER	SENSOR 1	SENSOR 2	SENSOR 3	SENSOR 4	PRINTER				
PCA 2 225	24-8350	02	CO							
PCA 2 235	24-8351	02	CO	NO						
PCA 2 245	24-8352	02	CO	CO (high)						
PCA 2 255	24-8353	02	CO	S02						
PCA 2 265	24-8354	02	CO	NO	NO2					
PCA 2 275	24-8355	02	CO	NO	S02					
PCA 2 225 Kit	24-8370	02	CO			Х				
PCA 2 235 Kit	24-8371	02	CO	NO		Х				
PCA 2 245 Kit	24-8372	02	CO	CO (high)		Х				
PCA 2 255 Kit	24-8373	02	CO	S02		Х				
PCA 2 265 Kit	24-8374	02	CO	NO	N02	Х				
PCA 2 275 kit	24-8375	02	CO	NO	S02	Х				

### ISO 9001

Printed in U.S.A. Prin. Tech

PCA<sup>®</sup> and Bacharach<sup>®</sup> are registered trademarks of Bacharach Inc. <sup>®</sup>May, 2007, Bacharach, Inc., all rights reserved. All information herein is subject to verification Product Bulletin - 8002 06/07 2M

**Technical Training for Heating & Cooling** 

All instruments can be upgraded to include combination of CO (high), NO, NO, and SO,

degree of measurement accuracy.

**Distributed By:** 

\* The Compact Sample Conditioner is recommended when measuring NO, and SO, to ensure the highest

Headquarters: 621 Hunt Valley Circle New Kensington, PA 15068 U.S.A Website: www.bacharach-inc.com

Phone: 1-800-736-4666